



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER OF PATENTS AND TRADEMARKS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/702,513	10/31/2000	Michael J. Leveille	19921/45	9261

7590 05/07/2003
Anthony J Janiuk Esq
Waters Corporation
34 Maple Street
Milford, MA 01757

EXAMINER

ROSENBERGER, RICHARD A

ART UNIT PAPER NUMBER

2877

DATE MAILED: 05/07/2003

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Applicati n N .

09/702,513

Applicant(s)

LEVEILLE ET AL.

Examiner

Richard A Rosenberger

Art Unit

2877

-- The MAILING DATE of this communication appears on th cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133).
- Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 08 April 2003.
- 2a) ☐ This action is FINAL. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-14 and 22 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☐ Claim(s) 1-14,22 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
- 11) ☐ The proposed drawing correction filed on _____ is: a) ☐ approved b) ☐ disapproved by the Examiner.
If approved, corrected drawings are required in reply to this Office action.
- 12) ☐ The oath or declaration is objected to by the Examiner.

Priority under 35 U.S.C. §§ 119 and 120

- 13) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. _____.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
* See the attached detailed Office action for a list of the certified copies not received.
- 14) ☐ Acknowledgment is made of a claim for domestic priority under 35 U.S.C. § 119(e) (to a provisional application).
a) ☐ The translation of the foreign language provisional application has been received.
- 15) ☐ Acknowledgment is made of a claim for domestic priority under 35 U.S.C. §§ 120 and/or 121.

Attachment(s)

- 1) ☐ Notice of References Cited (PTO-892) 4) ☐ Interview Summary (PTO-413) Paper No(s). _____
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948) 5) ☐ Notice of Informal Patent Application (PTO-152)
- 3) ☐ Information Disclosure Statement(s) (PTO-1449) Paper No(s) _____ 6) ☐ Other:

1. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

2. Claims 1-14 and 17 are rejected under 35 U.S.C. 103(a) as being unpatentable over Wynn (US 5,905,271).

Wynn shows a cell body having two ends and a fluidic channel allowing the passage of fluids (column 2, lines 6-9). There is an element holder (either of the "aligned apertures" 26, 27); the element holder has a substantially planar sealing surface (the bottom walls 31 of the recessed areas, which are sealing surfaces and are substantially planar). There is a stepped element (either of 16, 17) with a stem (19) and a base (21), the stem having an end surface and the base a substantially planar sealing surface. The stepped element is, in use, contained within the element holder and is sealed within the cell body by a sealing gasket (O-ring 29) positioned between the sealing surfaces of the element holder and the element. There are screws (37) which hold the structure together, exerting pressure on the gasket (31) between the two sealing surfaces. The stem portions of the elements create a measurement path length (column 2, lines 23-25).

The independent claims 1 and 17 refer to the pathlength created by the claimed structure as "non-adjustable". Wynn teaches the use of set screws (44) to "lock" the "adjusted bodies" (36) in place (see column 3, lines 3-5); once the "adjuster bodies" are so locked in place the pathlength is "fixed" and "non-adjustable".

Wynn discloses the use of screws (37) to hold the structure together. The use of bolts would have been an obvious substitution as the two are recognized equivalents.

It would have been obvious to make the windows out of any appropriate transparent material. It is known in the art to include lenses, either separately or by forming curved surfaces in the window, in sample cells; doing so in the device of Wynn et al would have been obvious. Those in the art could make design changes to the window and arrangement of Wynn such as varying the shape of the window while maintaining the functional aspects of the window and cell structure.

3. Claims 1-14 and 22 are rejected under 35 U.S.C. 103(a) as being unpatentable over Dätwyler et al (US 5,003,174) in view of Wynn (US 5,905,271) and Goldsmith (US 4,580,901).

Dätwyler et al shows a flow cell with a cell body and a stepped window (4,5) and a sealing gasket (32) which lies between a sealing surface on the window (7) and on the cell body (10). The sealing surface on the cell body of Dätwyler et al is not "substantially planar", but rather conical.

It is known in the art that stepped shaped windows of the type shown by Dätwyler et al can be sealed in a measuring cell with a sealing gasket when the sealing surface on the cell body is in a configuration other than the conical shape shown by Dätwyler et al. Wynn, for instance, shows substantially planar bottom walls (31) to seal the window.

It is also more generally known in the art that sealing gaskets can be used against planar surfaces to seal windows in fluid-containing sample cells. Goldsmith, for example, shows gaskets (23, 24) to seal a window ("radiant energy transmissive top 11") and a bottom element (13) to form a fluid-type cell structure.

Those in the art would have recognized, particularly given the known differences in construction among the sealing surfaces known in the art as illustrated by the references, that the exact shape of the sealing surface on the cell body is not critical; what is critical is that the sealing gasket fit against the window and the sealing surface of the cell body, whatever its shape, to seal the window so the cell does not leak. Thus those in the art would have found it obvious at the time the invention was made to use a "substantially planar" sealing surface in the place of the conical sealing surface of Dätwyler et al because it is, and would have been recognized as being, the function of sealing, and not the exact shape of the sealing surface, which is of importance in such situations, and would have been recognized that a "substantially planar" sealing surface would have accomplished this sealing function, while being a simple-to-manufacture configuration.

The use of bolts to hold the cell together would have been an obvious as such uses of bolts is well-known in the art; Goldsmith uses bolts (21) for such a purpose (see column 2, line 68), and also see the use of the similar screws in Wynn.


It would have been obvious to make the windows out of any appropriate transparent material. It is known in the art to include lenses, either separately of by forming curved surfaces in the window, in sample cells; doing so in the device of Dätwyler et al would have been obvious. Those in the art could make design changes to the window and arrangement of Dätwyler et aln such as varying the shape of the window while maintaining the functional aspects of the window and cell structure.

4. Papers related to this application may be submitted to Group 2800 by facsimile transmission. The faxing of such papers must conform to the notice published in the Official Gazette, 1096 OG 30 (15 November 1989). The fax number is (703) 308-7722.

5. Any inquiry concerning this communication or earlier communications from the examiner should be directed to R. A. Rosenberger whose telephone number is (703) 308-4804.

Any inquiry of a general nature or relating to the status of this application should be directed to the Group receptionist whose telephone number is (703) 308-0956.

R. A. Rosenberger
2 May 2003



Richard A. Rosenberger
Primary Examiner